

Operating Manual
NOAA/ESRL/GMD Aerosol Sampling System
Last Updated – February 4, 2010
Table of Contents

1.	General description	2
1.0.	Maintenance Checklist.....	4
1.0.1.	Daily Tasks	4
1.0.2.	Weekly Tasks.....	4
1.0.3.	Monthly Tasks	4
1.1.	GMD Contact information.....	4
2.	Maintenance.....	5
2.1.	Check System Parameters and Logging Status - (Daily Task).....	5
2.2.	CN counter butanol level - (Daily Task).....	5
2.3.	Changing the PSAP Filter - (Daily Task).....	6
2.4.	Check water level in humidifier reservoir - (Daily Task).....	8
2.5.	CCN Maintenance – supply and drain water bottles - (Daily Task).....	8
2.6.	Nephelometer Span Check - (Weekly Task)	9
2.6.1.	Span Check Plumbing.....	9
2.6.2.	Span Check Operation	10
2.7.	Servicing impactors for nephelometer and PSAP - (Weekly Task)	10
2.7.1.	Removal of Impactors.....	10
2.7.2.	Cleaning of Impactors.....	11
2.7.3.	Re-assembly of impactor box	12
2.8.	System Leak Test - (Weekly/Whenever impactors are serviced).....	12
2.9.	Change Carousel Filters - (Weekly Task).....	14
2.9.1.	Preparation	14
2.9.2.	Change the impactor films	15
2.9.3.	Changing the filters.....	15
2.10.	CCN maintenance - desiccant cartridge check (weekly task).....	16
2.11.	CN Butanol flush (monthly task).....	16
2.12.	Changing Nephelometer Lamp – (As needed, ~3-6 months)	17
2.13.	Resetting Love PID controllers – (As needed)	17
2.13.1.	Changing the setpoint	18
2.13.2.	Resetting	18
2.14.	System Clock	20
2.15.	Instant Messenger	20
3.	Computer Operation.....	20
3.0.	General.....	20
3.1.	Startup.....	21
3.2.	Shutdown	22
3.3.	Archiving and Backing-up Data	22
3.3.1.	Saving current data	23
3.4.	Other desktop icons.....	23
3.5.	LiveCPD Documentation.....	23
3.6.	Graphical Client Program (cpx).....	24

3.7.	Aerosol Data Logger Program	24
3.7.1.	General Operation	24
3.7.2.	Display	25
3.7.3.	Root Menu <enter> or <esc>	25
3.7.4.	Aerosol Status Window and Monitor Menu <enter><A><M>	27
3.7.5.	Nephelometer Status Window and Menu <enter><N><M> or <enter><O><M>	27
3.7.6.	PSAP Menu <enter> <L> <M> or <enter><R> <M>	28
3.7.7.	Filters Menu <enter> <F> <M>	28
3.7.8.	Humidograph Menu <enter> <H><M>	29
3.7.9.	CCN Menu <enter> <D><M>	29
4.	Safety	30
4.1.	Heater Control	30
4.2.	Electrical Safety	30
5.	Specifications	30
5.1.	Power Draw at 115 V DC	30
6.	Additional Computer info	31
6.0.	Data flow	31
6.1.	Aerosol Server Program (cpd)	31
6.2.	User account	31
6.3.	Location of this document	31
7.	Station specific appendix	32

1. GENERAL DESCRIPTION

The GMD Aerosol Sampling System is deployed at a number of stations around the world (see Table 1). It is an automated system for measuring aerosol light scattering as a function of wavelength, aerosol light absorption, and condensation nuclei. At several stations the light scattering as a function of relative humidity is also measured. 1 and 10 μm impactors are used to measure these properties for aerosol less than 1 and 10 μm , respectively. The main components are a three wavelength nephelometer (TSI model 3563), a light absorption photometer (Radiance Research PSAP) and a Condensation Nucleus Counter (TSI mode 3760). At some stations filter measurements are also made in conjunction with PMEL of the ionic composition of the aerosol. sea salt, non-sea salt (nss) sulfate, methanesulfonate, NH_4^+ , nss K^+ , Mg^{+2} , and Ca^+ for sub-micron and super-micron size ranges. Table 1 indicates which components are deployed at each station.

Table 1 Long term station list, locations and system components

Stn ID	Station Location	Neph	PSAP	CNC	Chem filters	Humid-ograph	other
ALT	Alert, Canada	X	X	X			
AMF	ARM Mob. Fac.	X	X	X		X	CCN
AMY	Anmyeon-do, Korea	X	X	X			
APP	Appalachian State	X	X	X			
BND	Bondville, IL	X	X	X	X		
BRW	Barrow, AK	X	X	X	X	X	CCN, SMPS
CPT	Cape Point	X	X	X			
CPR	Puerto Rico	X	X	X			
LLN	Lulin, Taiwan	X	X	X			
MLO	Mauna Loa, HI	X	X	X			
SGP	Lamont, OK	X	X	X		X	CCN, SMPS
SMO	Am. Samoa			X			
SPO	South Pole	X		X			
THD	Trin. Head, CA	X	X	X			
WHI	Whistler, Canada	X	X	X			
WLG	Waliguan, China	X	X	X			



Basic rack system (MLO)

A basic system generates the first four pieces of information; an extended system with all the options generates the following information:

- (1) Total and backwards scattering coefficients time series at 450, 550 and 700 nm wavelengths in two size ranges
- (2) Total and backwards scattering coefficients time series at 450, 550 and 700 nm wavelengths in two size ranges as a function of relative humidity
- (3) Absorption coefficient time series at 565 nm OR absorption coefficient time series at 467, 530 and 660 nm in two size ranges
- (4) Condensation nuclei concentration time series
- (5) Mass concentration of sea salt, non-sea salt (nss) sulfate, methanesulfonate, NH_4^+ , nss K^+ , Mg^{+2} , and Ca^+ for sub-micron and super-micron size ranges
- (6) Cloud condensation nuclei concentration as a function of up to 7 super saturations
- (7) Size distribution (diameter range depends on instrument)

From the measurements the following parameters can be calculated (1-3 for basic system):

- (1) Single scattering albedo at 1 OR 3 wavelengths in two size ranges
- (2) Angstrom coefficient
- (3) Hemispheric backwards scattering fraction in two size ranges
- (4) $f(\text{rh})$ if a humidograph system is installed.
- (5) CCN fraction (CCN/CN) if a CCN is installed

1.0. Maintenance Checklist

The system is mostly automated but there are some tasks that need to be performed on a daily, weekly or monthly basis. The tasks for a complete system are listed below and are described in more detail in **Section 2 – Maintenance**. More specific maintenance tasks for individual stations are described in each station’s appendix. Each station also has a binder of instrument manuals for each of the instruments and other system components to aid in more complicated trouble-shooting and maintenance (in consultation with station scientists).

1.0.1. Daily Tasks

1. Fill out daily checklist
2. Check logging status: letters at bottom of screen should be appropriate for that station (see appendix)
3. Check butanol level in CN counter and add butanol if needed
4. Change PSAP filter if transmittance < 0.9
5. Add distilled water to humidifier rack reservoir (if less than half full)
6. Add water to CCN supply bottle and drain the drain bottle

1.0.2. Weekly Tasks

1. Perform daily tasks
2. Change PSAP reference filter
3. Service nephelometer/PSAP impactors
4. Perform leak test (if weather permits)
5. Perform nephelometer span check (**Don’t** do during nephelometer/PSAP impactor servicing)
6. Change filter carousel and carousel impactor film (can be done during span check)
7. Mail carousel filter samples to NOAA/PMEL
8. CCN desiccant cartridge check

1.0.3. Monthly Tasks

1. Mail PSAP filters and daily check lists to NOAA/ESRL/GMD
2. Check supply levels of PSAP filters, plastic bags, butanol, ethanol.
3. CN counter butanol flush.
4. Check integrity of external tubing, fittings, insulation, etc.

1.1. GMD Contact information

Betsy Andrews Email: betsy.andrews@noaa.gov Office Phone: 303-497-5171 Home Phone : 303-442-5142	John Ogren Email: john.a.ogren@noaa.gov Office Phone: 303-497-6210 Home Phone: 303-499-4079
Anne Jefferson Email: anne.jefferson@noaa.gov Office Phone: 303-497-6493 Home Phone: 303-579-9013	Pat Sheridan Email: patrick.sheridan@noaa.gov Office Phone: 303-497-6672 Home Phone: 303-888-6602

2. MAINTENANCE

This section describes the general maintenance/station tasks for a station with all the options. Not all tasks are applicable to each station, so each task includes a list of stations where it must be performed.

2.1. Check System Parameters and Logging Status - (Daily Task)

Stations applicable: ALT, AMF, AMY, APP, BND, BRW, CPT, CPR, LLN, MLO, SGP, SMO, SPO, THD, WLG

Time required: 5 minutes

Tools/supplies: Daily check list

Overview: There are a few system parameters that are not recorded and hence need to be checked manually. A daily checklist is used to record the values and is mailed back to GMD at the end of each month with the used PSAP filters.

1. Fill out daily checklist (and weekly/monthly on back of sheet if applicable)
2. Check that letters on bottom of computer screen match the letters listed in station specific appendix/checklist.

2.2. CN counter butanol level - (Daily Task)



Stations applicable: ALT, AMF, AMY, APP, BND, BRW, CPR, LLN, MLO, SGP, SMO, SPO, THD, WHI, WLG

Time required: 5 minutes

Tools/supplies: butanol fill bottle

Overview: The CN counter condenses butanol vapor onto particles in the sample line to grow them to a size so that they can be seen and counted by the instrument optics. The condensed butanol goes out the exhaust stream to the butanol trap and fresh butanol must be added to the counter to replenish the supply.

Check the butanol level – it should be about 1/2 way up the **view port**. **Do not overfill.**

	
This is the new style CN box with a CN model#3010 in it.	This is a CN model#3760. It is filled the same way as the model#3010

If butanol needs to be added connect the butanol fill bottle to the butanol **fill port**. Loosen the cap on the butanol bottle and fill the CN counter until the butanol level shown in the view port is acceptable.

2.3. Changing the PSAP Filter - (Daily Task)

Stations applicable: ALT, AMF, AMY, APP, BND, BRW, CPT, CPR, LLN, MLO, SGP, THD, WHI, WLG

Time required: 5 minutes

Tools/supplies: PSAP Kit


Overview: The PSAP filter transmittance degrades as particles collect on it. To ensure valid measurements the filters need to be changed.

Is the PSAP filter transmittance above 0.9? (Look at Tr value, in picture below Tr= 0.495)




If not, the filter needs to be changed.

1. In the Aerosol Data Logger window, press <Enter><L><M><2> or <enter><R><M><2> to select the PSAP menu and signal start of a filter change. The bottom of the PSAP window will indicate that a filter change is taking place.



2. Unscrew the knob on the sampling head, raise the lockdown panel, and remove the filter holder.



3. Remove the filters. **It is only necessary to replace the reference filter once per week.** Place the **sample** filter (the left-hand one) in a small zip bag and label the bag with the station name and the date the filter was pulled out of the PSAP. The label should have the format: YYMMDD-THD (so December 13, 2000 would be labeled: 001213-THD). Write your initials on the bag.

4. Place new filters in the filter holder. The white side of the filter should be facing up. Re-insert the filter holder into the sampling head; secure the lockdown panel with the black knob. Make sure the lower lip of the lockdown panel goes under the bottom edge of the sampling head.

5. Set PSAP flow to the following (in lpm): BRW=2; MLO=3; THD=1; BND=0.5; SGP=0.75; CPR = 1.0, APP=1.0, AMY=0.5 using the knob on the right side of the PSAP.



6. On the PSAP, press the “RESET Transmittance” switch up and verify that $Tr=1.000$ on the display of the PSAP.

7. If the PSAP window still says ‘Changing filter...’ then, tell the system you are done by pressing `<Enter> <L><M><2>` or `<Enter><R><M><2>` to select the PSAP menu and signal end of the filter change. Verify Transmittance = 1.000 before answering the “Y” when asked, “Have you pressed the RESET Transmittance switch?” Note: If the PSAP window no longer says ‘Changing filter...’ then the system has already determined you are done with the filter change. If the system is frequently determining the filter change is over before you’ve told it the filter change is over the ‘Threshold_RSD’ setting in the cpd.ini file may be incorrect. Check in with

your NOAA contact (John Ogren or the scientist assigned to your station) to find out how to deal with this.

2.4. Check water level in humidifier reservoir - (Daily Task)

Stations applicable: AMF, BRW, SGP



Time required: 5 minutes

Tools/supplies: distilled water

Overview: The humidifier condenses water vapor onto particles after they exit the reference nephelometer, but before they enter the humidified nephelometer. The condensed water vapor exits the system and must be replenished.

The water reservoir needs to be refilled whenever the level in the reservoir gets below ~1/2. *If reservoir has lost lots of water in a short period of time this may indicate a leak! Inspect for water around the system or for an RH value in the second nephelometer above 90%. If there is a leak a) turn off humidifier heater and preheater b) unplug the water pump c) enter a log message in the computer <Control><M><enter>reservoir leak!! <enter>*

Lift the attached fill funnel above the top of the reservoir bottle. Slowly pour distilled water into the funnel. Do not fill reservoir above the bottom of the red tape line.

	<p>Humidifiers have square box for water reservoir- see yellow arrow.</p>		<p>Note at some stations you may be able to merely lift the lid and pour the water into the water reservoir. At others you may need a funnel and tubing to conveniently add water.</p>
--	---	---	--

2.5. CCN Maintenance – supply and drain water bottles - (Daily Task)


Stations applicable: AMF, BRW, SGP

Time required: 10 minutes daily

Tools/supplies: distilled water

Overview: The CCN instrument creates supersaturated conditions inside its main column to activate particles into droplets, mimicking the processes that form clouds. To create these supersaturated conditions the main column must have a continuous supply of water coating its sides. **Always drain the drain bottle when filling the supply bottle!**

Open the door on the lower left of the CCN instrument (figure a). Check the water level in the supply bottle (the front bottle) (figure b). If the water level is lower than 100 mL then fill the supply bottle needs to be filled and **the drain bottle needs to be drained**. To remove bottles, first undo the long wingnut screw holding the collar on the bottle necks (figure c). Next hold the cap of the supply bottle in place (so as not to mangle the tubing!) and twist the bottle clockwise. Fill the supply bottle with distilled water to the fill line. Remove and empty the water from the drain bottle. Replace the drain bottle and then the supply bottle and tighten screw. If the drain bottle is not drained when the supply bottle is filled there is a risk of flooding the CCN optics.



2.6. Nephelometer Span Check - (Weekly Task)

Stations applicable: ALT, AMF, AMY, APP, CPT, CPR, BND, BRM, BRW, LLN, MLO, SGP, SPO, THD, WHI, WLG

Time required: Approximately 40 minutes

Tools/supplies: None

Overview: The nephelometer span check confirms that the nephelometer calibration is correct. The procedure is mostly automated, although the valve on the CO₂ tank needs to be manually opened and closed. The system control code CPD automatically turns off the heaters on the systems that have humidographs (BRW, SGP, AMF), so that humidified air does not contaminate the span check.

2.6.1. Span Check Plumbing

The plumbing for a span check is as follows: CO₂ tank → Regulator → Flow control Valve → Rotameter → Copper tubing → Plastic tubing → filter → neph. CO₂ inlet (wet neph inlet at sites with humidograph (AMF, BRW, SGP))



2.6.2. Span Check Operation

Open CO₂ tank valve. At stations with multiple nephelometers (typically stations with humidograph systems (e.g., BRW, SGP, AMF)) the CO₂ flow rate should be ~8lpm, at all other stations the CO₂ flow rate should be ~5lpm.

Press <enter> <N><M><2> to begin span checks (this will start span checks and CO₂ flow through either the single nephelometer or both nephelometers). Note: for stations with multiple nephelometers you could instead press <enter> <O><M><2>.

The computer will prompt you for CO₂ tank pressure and flow rate and will notify you when the span check is over and display the calibration results. You will need to toggle between nephelometer windows to check both nephelometer's span check results. Typing these sequences: <enter> <N> and <enter><O> a couple times each will take you to each nephelometer's window and the corresponding span check results.

Conduct a second span check if the average deviation is more than 4% for either nephelometer. The point of the second span check is to make sure that the large error in the first span check wasn't caused by operator error (e.g., forgetting to open the CO₂ valve or not having the CO₂ valve open enough). **If the second span check is also higher than 4% then contact a NOAA scientist to discuss troubleshooting the problem.**

Close CO₂ tank valve when finished.

2.7. Servicing impactors for nephelometer and PSAP - (Weekly Task)

Stations applicable: ALT, AMF, AMY, APP, BRW, CPT, CPR, LLN, MLO, THD, BND, SGP


Time required: 20 minutes

Tools/supplies: a 1 1/8" wrench, and the impactor kit

Overview: It is necessary to clean the impactor plate to ensure the impactor size cut is accurate. Cleaning the impactor plate involves disconnecting the impactors and disassembling them so that the impaction place can be accessed and cleaned. *Cleaning should occur more frequently than weekly if large particle deposits are observed.*

2.7.1. Removal of Impactors

1. <enter><enter> Enter message "Starting impactor servicing". <enter>
2. <enter><A ><M><2> to select "bypass analyzers". This closes a solenoid valve and notifies the computer that impactors are off-line. Close the **Manual ball valve**. This is the valve upstream (behind) the 10 um impactor in the impactor box

	<p>This is a top view of the impactor box.</p> <p>Manual ball valve is in upper left of picture (hard to see).</p>	<p>Close the manual ball valve behind the 10 um impactor by rotating 90 degrees to right (clockwise).</p>
---	--	--


3. Using 1 1/8" wrench, loosen the appropriate fittings and carefully remove impactors. Be careful of the tubing and cables in the impactor area so as not to disturb them.
4. Place impactors on a flat clean surface. Use a large clean Kimwipe or printer paper.

2.7.2. Cleaning of Impactors

HINT: If you lay aside the cleaned pieces in order of removal, it will be easier to reassemble afterwards.

1. Unscrew black knobs on 10-um impactor, and fold down the side arms.
2. Remove top piece, clean inside (including inlet) with alcohol, tissue, and cotton swabs.
3. Remove impactor jet plate, clean both sides with alcohol, tissue, and cotton swabs.
4. Remove spacer ring, wipe clean.
5. Remove impaction plate, clean both sides with alcohol, tissue, and cotton swabs.
6. Remove bottom spacer ring that impaction plate sits in, wipe clean.
7. Clean inside of outlet fitting.
8. Replace bottom spacer ring.
9. Spray an impaction plate with a thin coat of silicone lubricant spray. Wipe the excess silicone spray off the vertical edges of the plate
10. Replace impaction plate.
11. Replace spacer ring.
12. Apply a generous coating of silicone vacuum grease to the impaction cup on the top surface of the impactor jet plate (only on 10-um impactor).
13. Replace impactor jet plate.
14. Replace top cover
15. Tighten assembly back together with black knobs

Sketch of impactor



2.7.3. Re-assembly of impactor box

1. Reinstall impactors in system being careful not to strip threads on fittings. Tighten all fittings by hand and then again with 1 1/8" wrench.
2. Open Whitey valve by rotating handle 90 degrees to right (so it is parallel with flow).
3. `<enter> <A><M><2>` again when the impactors have been re-installed, to “unbypass analyzers” and to resume data logging. The ball valve message is asking if you have opened the Whitey valve. Enter ‘Y’ once the valve is open.
4. `<enter> <N><M><4>` to start a nephelometer zero cycle
5. Enter a log message: `<enter><Enter>` "System back on-line after impactor servicing and zeroing." `<enter>`
6. Do a leak check (see below, weather permitting) to make sure all fittings have been tightened and no leaks have been introduced into the system during the impactor change.

2.8. System Leak Test - (Weekly/Whenever impactors are serviced)

Stations applicable: AMF, AMY, APP, BND, CPT, CPR, LLN, MLO, THD

Time required: Approximately 10 minutes

Tools/supplies: Large white HEPA filter with 2" cup fitting


Overview: It is important to test the aerosol system for leaks to ensure that measurements are made on outside/conditioned air and not on inside air. Two diagnostics are used to test the aerosol system for leaks: the pressure drop and the CN counts. Note this is easier/quicker with 2 people – one person to place and remove the filter at the saddle port and one person to enter the observed values for pressure drop and CN counts.

DO NOT attempt this procedure during stormy weather; SAFETY FIRST!! ☺

Enter a log message <enter><enter> "Starting leak check" <enter>.

Bring up the aerosol monitors window <enter><A>.


Turn 3-way valve at CN inlet to 'Neph' position (i.e., it is perpendicular to flow into the CN counter inlet). Note: don't do the leak check between XX:57 and XX:03 (3 minutes before the hour to 3 minutes after the hour) during that time the neph is zero-ing automatically and you will measure 0 particles when the valve is pointed towards the neph.



Back of CN box. Photo shows 3-way leak check valve in 'normal' position. For a leak check, rotate the 3-way valve clockwise 90° towards neph (as indicated by yellow arrow)


Note: don't do the leak check between XX:57 and XX:03 (3 minutes before the hour to 3 minutes after the hour). During that time the neph is zero-ing automatically and you will measure 0 particles when the valve is pointed towards the neph.

Enter a log message <enter><enter> "Ambient CN count= value; dp_Neph_imp= value" <enter>. The CN counts are the value for N_cncc in the main cpdclient window, or the CN counts can also be found by opening the CN window: <enter><c>. The dp_neph_imp value is in the center right of the aerosol window.



The value you would enter for Ambient CN count is outlined in pink in this picture. The value you would enter for dp_Neph_imp is outlined in yellow. Based on this picture, the message you would enter is: Ambient CN count=1603; dp_Neph_imp=0.0

Climb up to the saddle port at the base of the 8" PVC, undo the pins and remove the saddle port. Place the filter with the metal cup securely on top of the exposed 2" metal inlet stack.



Enter a log message `<enter><enter> "Filtered CN count= value; dp_Neph_imp= value" <enter>`.

Acceptable ranges for these two variables when the filter is on the 2" sample inlet are:

`dp_neph_imp > 20` `CN_counts < 5% of ambient value`

If the observed values are outside these ranges, there may be a leak in the system. A likely location to check is the connections on the impactors. If tightening those fittings doesn't help, call the GMD contact person.

Remove the filter and replace saddle on 8" PVC.

Return the 3-way valve to Inlet position (i.e., parallel to CN inlet sample).

2.9. Change Carousel Filters - (Weekly Task)

Stations applicable: BRW, BND

Time required: 30 minutes

Tools/supplies: Kimwipes, new canister of filters, old canister of filters, glovebox, 1 1/8" wrench, silicon spray.

Overview: Aerosol filters need to be removed from the carousel weekly and sent in to be analyzed.

2.9.1. Preparation

1. Prepare glovebox by putting down new Kimwipes and a new pair of outer gloves. Every other week or when you notice holes in the vet gloves replace them (you will need to tape these gloves on).
2. Turn on glove box pump.
3. Prepare a clean flat surface for changing the impactor films.
4. Type '`<enter><F>,<M>,<5>`' to select filter change and notify the computer that the filters are off line.
- 5a. Close manual **Whitey ball valve**, which is above filter carousel, behind impactor by moving it 90 degrees clockwise so that it is perpendicular to the flow.



- 5b. Select option #2 to inform cpd that the ball valve is closed.
6. Release 8 white tubes with the quick connect/disconnect fittings, and disconnect the fitting at the elbow on the black tubing connecting the impactor to the carousel.
7. Place the carousel and a canister of filters (write start date on outside) in the glovebox, close the glove box and turn on the pump.
8. Back at the rack, disconnect fitting near valve so the impactor can be removed.

2.9.2. Change the impactor films

1. Wear gloves
2. Unscrew the black knobs and remove outer white sleeve. Make a note of the order of each section so that they can be put back in the correct order.
3. top section: clean out grease using methanol and Kimwipe; apply new grease
4. Section 2 (>10 μm particles): throw out film; clean section; spray a film from the plastic bag with silicon lubricant and place in holder with spacer ring on top
5. Section 3 (1-10 μm particles): take out film with tweezers and place in old 'S' Petri dish. Clean section; Place new 'B' film on the stage, let sit for a minute and then return to 'B' Petri dish. This is the new 'background' film. Put new 'S' film on stage, place spacer ring on top and put impactor back together.
6. Place impactor back in rack.

2.9.3. Changing the filters

1. Take lid off carousel; black knobs first, then white knobs, turning knobs in counter-clockwise direction. (If you have difficulty doing this, try doing this outside of the glove box but put a note in the filter box when you do.)
2. Take out the red discs from the old container. Remove filters from carousel and stack in the old container in reverse numerical order (8 to 1), putting a red disc in between each filter. Put on the container cap and screw down.
3. Clean the carousel with alcohol. Place the new filter samples in the carousel, matching up the filter number with the carousel number.
4. Put top back on carousel, turn white knobs then black knobs
5. Put carousel back in the rack, attach the quick connects (orange-to-orange and black-to-black). Attach black hose to impactor.
6. Open Whitey valve by rotating handle 90 degrees clockwise so it is parallel to the flow.
7. Do a leak check (weather permitting) to make sure all fittings have been tightened
8. Type '`<enter><F>,<M>,<5>`' to end filter change and begin filter sampling, select option #2 to inform computer that the ball valve is open..
9. Record stop date on old sample container, put in box along with Petri dishes to mail to PMEL:

Kristen Schulz
 NOAA/PMEL
 7600 Sand Point Way NE
 Seattle, WA 98115

phone: 206-526-6220
 email: Kristen.Schulz@noaa.gov

2.10. CCN maintenance - desiccant cartridge check (weekly task)


Stations applicable: AMF, BRW, SGP

Time required: 30 min

Tools/supplies: 7/64" hex wrench, Phillips or straight screwdriver, spare desiccant cartridge

Overview: The CCN counter counts the number of droplets created in its main column using an optical particle counter (OPC). To protect the OPC optics from the wet environment the droplets are in, a dry air flow is passed over the OPC optics window. Occasionally (~2X/year) the desiccant cartridge which generates the dry air must be replaced.

To check the status of the desiccant cartridge, remove the CCN cover counter. First remove the computer screen –slide it up ~1/2" and then pull off and place somewhere (you don't need to undo the cables). Move the keyboard out of the way. To remove the cover you will need a screwdriver to undo the screws holding the cover to the back of the instrument (figure a), there are ~10 screws: 3 on top, 4-5 on the left, 2 on the right. Next you will need to undo 3 hex nuts on the left side of the cover (figure b) using the 7/64" hex wrench. Finally, pull the cover out of the rack towards you so you can inspect the desiccant cartridge (figure c). The cartridge should show blue above the label at the bottom. If there is no blue showing (you only see pinkish-purplish colors) the cartridge should be replaced.



2.11. CN Butanol flush (monthly task)

Stations applicable: AMY, CPR, THD (at CPR it's a weekly thing!)

Time required: 15 min

Tools/supplies: 9/16" wrench Phillips, waste butanol drain bottle

Overview: The butanol in the CN counter absorbs water readily, so at humid sites the butanol should be flushed monthly to maintain the appropriate butanol properties and to keep the optics block clean.


Enter a log message: <enter><Enter> "Doing a butanol flush." <enter>

Detach the CN counter vacuum so no flow is going through the counter (at THD the vacuum line is accessed from the back of the rack – it is labeled CN vacuum). Attach the drain bottle to the CN drain port on the front of the rack and loosen the cap. Butanol should flow into the drain bottle. Note: if butanol does not seem to want to flow, (1) detach the drain bottle (2) squeeze the bottle with the cap loose (3) tighten the cap keeping the bottle squeezed (4) reattach the drain bottle to the drain port (5) loosen the cap. This procedure should create a small suction to the bottle and get the butanol flow started.


Inspect the butanol you've drained out. If it looks pretty clear and not yellowish, refill the CN counter with fresh butanol and reattach the vacuum line. If the butanol is not clear and/or has a yellowish tinge, refill and drain the CN counter several times to try to clean the gunk out of the instrument. Contact the station scientist if the butanol remains discolored despite multiple drains.

2.12. Changing Nephelometer Lamp – (As needed, ~3-6 months)

The system mentor will let you know if this needs to be done. The nephelometer lamp usually last about six months. Typically, scientists watching the station data (lamp voltage and current) at GMD can tell about a week ahead of time when the lamp is going to go. We like to change the lamp before it goes due to the possibility that when the lamp goes it could cause further damage to the nephelometer. Detailed instructions for changing the nephelometer lamp are given in the TSI nephelometer manual on page 8-13.





This is the neph window in the data acquisition program. You can see the lamp voltage (Lamp V) and current (Lamp I) at the lower right.



This is a plot of the neph lamp voltage (green) and current (red) as the bulb started to die at one of our ground based stations. It's not instantaneous so you should be able to catch it if you keep an eye on it.

2.13. Resetting Love PID controllers – (As needed)

Love PID controllers are used in the aerosol system to control heaters (based on a relative humidity reading) and to control flow for the main sample line(s).

	Front of PID box show 4 PID controllers
	<p>The top value (red numbers) is the measured value. The bottom value (green numbers) is the set point.</p> <p>The button on the far left is the INDEX button.</p> <p>The button on the far right is the ENTER button.</p>

2.13.1. Changing the setpoint

Typically the Love PID controllers require no maintenance or changes. However, at some point it may be necessary or useful to change a set point. Hit the **INDEX** button to get to the primary menu setting SP1. Use the up/down arrows (the two middle buttons) to adjust the setpoint to the desired value. Then hit **ENTER** to make the setpoint take effect.

2.13.2. Resetting

Occasionally (quite rarely) the Love PID controller will fail in a non-catastrophic way and need to be reset. If this happens, the factory default values should be restored and then the settings should be adjusted to those for the aerosol system. The procedure is described below.

This should be done if any strange messages (e.g., FAIL tEst) appear on the controller display)

- Turn off power to controller
- Turn on power to controller
- While controller is performing **SELF tEst**, press and hold the **INDEX** and **ENTER** keys.
- The controller will display the ROM ID code. Press **INDEX**.
- The controller will display **FAcT dFLt**. If you wish to just restore factory settings, press **ENTER** and **DOWN ARROW** at the same time. The controller will be reset to the original factory settings.
- Press **INDEX** to display **ACPt**. Select **YES** or **no**.
 - YES** -- Changes are accepted and control re-boots.
 - no** -- Changes are discarded and control re-boots.

Once the controller has gone through the reset procedure and has the factory default settings in it you will need to set the settings to those that are used for the aerosol system. The table below describes the settings used by the aerosol system. There are 3 different PID menus that need to be adjusted.

- (1) Primary Menu – access using the **INDEX** button

- (2) Secondary Menu – access by pressing the up arrow button and the **INDEX** button at the same time.
- (3) Secure Menu – access by pressing and holding the up arrow button and the **INDEX** button at the same time for 5 seconds

Once you are in a menu domain, use the **INDEX** button to move up or down; use the up/down arrows to change values and use the **ENTER** button to keep a changed value.

Settings for PID Controllers should be as listed in the table, unless otherwise noted. You may see some other parameters listed in your PID menus, but if they are not specifically listed below, they are not applicable and can be ignored. Pictures of a Love PID controlling showing each of the settings are in the document: /aer/doc/inst/pid/LovePID_settings.doc.

PID Menu settings for aerosol system

Love PIDs for heater control (#1,#2,#3)			Love PID for flow control (#4)		
Primary	Secondary	Secure	Primary	Secondary	Secure
Auto = On	ALLo = 0.0	SECr = 4	Auto = Off	Out1 = ProP	SECr = 4
SP1 = 40.0	ALHi = 100.0	InP = diFF	SP1 = 30.00	tun = nor	InP = Volt
	Out1 = 1tP	Unit = none		ArUP = OFF	OSUP = OFF
	Tun = nor	dPt = 0.0		ArtE = OFF	Unit = none
	ArUP = OFF	InPt = OFF		Fint = 0	dPt = 0.00
	ArtE = OFF	INPb = FAIL		Pct0 = OFF	InPt = OFF
	Fint = 0	APct = AdJ		Prog = OFF	InPb = FAIL
	Pct0 = OFF	SEnC = OFF		PSEt = OFF	APct = rEAL
	Prog = OFF	SCAL = -100.0		InPC = 0.00	SEnC = OFF
	PSEt = OFF	SCAH = 100.0		FiLt = 2	SCAL = 0.00
	InPC = 0.0	SPL = 0.0		LPbr = OFF	SCAH = 99.99
	Filt = 2	SPH = 100.0		LOrE = rE	SPL = 0.00
	LPbr = OFF	SP10 = OutA		Addr = 34	SPH = 50.00
	LOrE = rE	SISSt = dir			SP10 = OutA
	Addr = see note below	SIOL = 0			SiSt = rE
		SIOH = 100			SIOL = 0
		SILP = 0 on			SIOH = 100
		AL = HiLo			SILP = 0 on
		Alt = AbS			Addr = 34
		ALrE = OnOF			bAUd = 9600
		ALPi = OFF			nAt = OFF
		AL iH = OFF			Stor = no
		ALSt = OPEn			
		ALLP = OoFF			
		ALLb = OFF			
		Addr = see note below			
		bAUd = 9600			
		nAt = OFF			
		Stor = no			

Note: Addr for PID #1, #2 and #3 should be 31, 32 and 33 respectively.

2.14. System Clock

The clock is synched automatically using the GPS unit or the local time server.

Table 2 Difference between local standard time (daylight savings) and UTC for some stations.

STN	APP	BRW	BND	CPR	MLO	SPO	SMO	SPO	SGP	THD
hrs diff	5 (4)	9	6 (5)	4	10	0	11	0	6 (5)	8 (7)

2.15. Instant Messenger

A text based instant messenger system is available within CPD. This is useful for troubleshooting with one person at the field computer and one person in Boulder. To use the instant messenger, both the person in the field and the person back in Boulder need to be logged into the field computer and running cpdclient.



Then both people follow these directions:


Bring up the main aerosol system menu <M> or <enter>. Choose the option 'Instant Messaging' either by using the arrow keys or selecting <4> and hitting <enter>. Then choose <M> to get a box for typing in messages. Hit <enter> to send the typed message. When you are done with all the instant messaging, hit <W> to close the window.

3. COMPUTER OPERATION

3.0. General

The computer runs the RedHat 9 distribution of the Linux operating system, with the kernel upgraded to version 2.4.22. The data acquisition and instrument control is done with a family of programs, collectively referred to as 'cpd', which uses a client-server architecture to separate the continuous operating tasks (acquiring data, writing to disk, controlling instruments) from the user interface tasks. The server is called 'cpd'. There are currently two clients, 'cpdclient' for text-mode display (i.e., the screen with windows displaying values for different measurement parameters), and 'cpx' for graphical (stripchart) display. The text-mode client allows the user to issue commands to 'cpd'.

	
<p>cpdclient (text mode) display also called aerosol data logger window</p>	<p>graphical (stripchart) display (CPX)</p>



←cpd2 desktop – this is an upgraded version of the cpd server system. It is functionally the same as the cpd desktop but looks slightly different.

- Most icons now on desktop (rather than in bottom toolbar.)
- Help icon on bottom toolbar pops up menu with useful documents (e.g., this operations manual) and contact info
- All desktop icons now require double clicking to activate.

The system clock is set to UTC (6 hours different than Boulder time during daylight savings). The clock is synchronized with a network time server using the 'ntp' protocol. If a network time server is not available, it is possible to use a GPS receiver directly connected to the computer as the time standard. The system clock is displayed in the lower right corner of the Aerosol Data Logger window.

Instruments are connected to the computer via RS-232C ports.

3.1. Startup


This assumes you have a USB stick which is already initialized and a LiveCPD CD. The CD contains everything you need to run CPD on the computer, while the initialized USB stick contains station specific configuration information (e.g., the cpd.ini file) and, depending on configuration, the USB stick is usually the location where data is stored.

Insert USB stick in USB port (either on back of computer or in edgeport module)

Turn on PC.

After a few minutes, the NOAA aerosol data logger window will appear.

There are eight at the bottom left of the screen. The shutdown icon is the ninth icon at the very far right of the screen. Position the mouse over each one to see its function.



CPD desktop functions, left to right

CPD2 desktop functions, left to right

<p>These functions, from left to right, are:</p> <ol style="list-style-type: none"> 1. Start data logger – the CPD program 2. Stop data logger 3. Stripchart data – the CPX program 4. Send data to NOAA 5. Edit the cpd.ini file (using Nedit) 6. Web browser 7. Nedit – which brings up a simple text editor window 8. Icon to bring up a terminal window 9. Shutdown everything! 	<p>These functions, from left to right, are:</p> <ol style="list-style-type: none"> 1. Start data logger – the CPD program 2. Stripchart data – the CPX program 3. Send data to NOAA 4. Edit the cpd.ini file (using Nedit) 5. Web browser 6. Nedit – which brings up a simple text editor window 7. Icon to bring up a terminal window 8. Stop data logger 9. Shutdown everything!
click ONCE on the icon to activate	Click TWICE on the icon to activate

During normal operation you should never need to select 1 because the data logger starts up automatically when the computer is turned on. Number 4 also happens automatically, with a frequency between 1 and 4 times per day (an automatic email is generated at GMD detailing what data files have been processed. Occasionally there are some automated warnings based on system parameters in the email as well.

If for some reason you did need to start cpd and the data logger, here's the what to do:
 To start the data acquisition program, click on the "Start cpd" icon. This will start 'cpd' in a background window, and several seconds later starts 'cpdclient' in the terminal window.
 To start the graphical client 'cpx', click on the "Stripchart data" icon. This takes a little while to get started so don't keep clicking if it doesn't start right away!! To stop data acquisition and shut everything off, click on the "Shutdown" icon. The "Shutdown" icon will ask whether you want to shut everything down, reboot or cancel the request to shutdown.

These programs can also be started from a command line in a terminal window by typing one of the following commands:

- cpdclient - starts the text mode client
- cpd.start - starts both the server and text mode client
- cpd.kill - kills the server
- cpx- starts the graphical client (X-windows only)

3.2. Shutdown

Terminate the data acquisition software by clicking on the "Shutdown" icon.
 Click on the "Main Menu" icon, then click on the "Log out" icon.
 On the login screen, click on "System" and then "Halt".
 Wait for the computer to display "Power down".
 Turn off power.

3.3. Archiving and Backing-up Data

Note: This is typically done automatically, these steps are needed only if manual archiving is required. The data archives are standard "ZIP" format. The data files are plain ASCII, comma-delimited, with headers.

3.3.1. Saving current data

If up-to-the-minute data are desired, press <enter><2><2> in Root Menu to close current files and start new files.

3.4. Other desktop icons

In the left hand side of the desktop there are two icons.





Clicking on the CPD data icon opens a window with the file directory in which the current data is being logged. (Equivalent to /aer/stn/log).

If you need to transfer data or something from the field computer to a windows compatible USB stick, insert the USB stick into the USB port on the back of the computer or in the Edgeport device, and click once on the USB drive icon. This will bring up a window with the directory structure of the USB stick. (Note: you may need to refresh the directory window if a different USB stick had previously been inserted.) When you are done with transferring data (or whatever) remove the USB stick – the system will automatically dismount within 30 seconds or so. The USB icon has a little dot on it while the USB stick is mounted and the dot goes away when it automatically dismounts.

3.5. LiveCPD Documentation

There are documents on the desktop explaining how this all works.

 A screenshot of the LiveCPD desktop. In the upper right corner, there are two icons: 'INI Documentation' and 'Readme'. The 'INI Documentation' icon is a document icon with a blue and yellow color scheme. The 'Readme' icon is a document icon with a blue and white color scheme. Below these icons, there is a row of smaller icons for 'StartCPD', 'StopCPD', 'CHK', and 'Exp'.	 A screenshot of the LiveCPD2 desktop. A help menu is open, showing a list of links: 'INI Documentation', 'Operating Manual', 'LiveCPD2 Documentation', 'System Status', and 'Contacts'. The 'Help' icon is located in the bottom left corner of the desktop, and the time '19:03:24' is displayed in the bottom right corner.
<p>In cpd there are icons in the upper right hand corner of the screen that link to useful documents</p>	<p>In cpd2 the help icon on the left of the bottom toolbar will bring up a menu with links to useful documents</p>

The INI Documentation icon opens a text document describing how to set up a cpd.ini file. The README icon describes the LiveCPD application in gory detail – how to configure the USB stick, more detailed descriptions of all the icons, etc. This document is called ‘LiveCPD2 documentation’ in CPD2. The LiveCPD/liveCPD2 documentation also explains the use of the icons in the lower right of the desktop.

3.6. Graphical Client Program (cpx)

The graphical client is a work in progress. Please send suggestions for improvement to John.A.Ogren@noaa.gov. Documentation for cpx can be viewed by clicking on the “Help” item on the cpx menu bar. This graphical interface allows you to see what’s happening with the data over time, so you could use it to watch CN counts get much lower when you put the leak test filter on the stack.

3.7. Aerosol Data Logger Program

3.7.1. General Operation


Press <enter> to access menus, <Esc> to abort from any menu or entry box without any changes, a second <Enter> to accept the current entry or highlighted item. Any instrument window can be displayed by the shortcut key <Enter><X>, where “X” is the code letter for the desired instrument. Windows are closed by pressing the <esc> key. Pressing the <Tab> key cycles the view among all the active windows. Windows can be repositioned on the screen using the arrow keys. Menu entries can be selected with the arrow keys or pressing the numerical digit for the desired selection.

The program will request an access code if you select a menu item that will result in a change to the data being logged to disk, in order to prevent inadvertent changes. The access code is ‘cmdl’ (lowercase is necessary).

Keystrokes	Action
<enter>	Shows root menu
<esc>	Shows root menu and hides any active menu
<Enter><X>	Selects root menu entry ‘X’ when root menu is not focused (X can be A,B,C,D,F,G,H,L,M,N,O,S,U,V,W or others depending on the station)
<enter><2>	Disc operations menu
<enter><enter>	Online log entry (message window)
<F12>	Online log entry (message window)
<control><E>	Toggles red error window (if present)
<control><C>	Quit cpdclient
<M>	Shows menu for active window
<esc>	Closes active window
<tab>	Cycles through open windows
Arrow keys (in menu)	Move thru menu list
Number/letter (in menu)	Quick select on menu list
Keystrokes	Action
<enter> (in menu/dialog box)	Chooses or toggles selected item
<enter> (dialog box)	Begins enter input mode, text will be bolded; pressing <enter> again will end input mode
<esc>	Exits menu/dialog box without applying changes
<enter> (with bottom entry selected in dialog box)	Closes dialog box and applies all changes
<enter><enter>	Brings up message window

3.7.2. Display

The “base” display consists of a “root” window for displaying current values for ambient measurements and a status line at the bottom left of the screen.



← status line location (in pink box) in root window

← example status line

The bottom status line gives the code letters of the active instruments and the system date and time. If the code letter for an instrument is absent, then that instrument is not present or not responding.


The code letters are assigned in cpd.ini, and vary from station to station. Status windows for specific instruments can be displayed by pressing <enter> plus the code letter for that instrument. Note: the code letters are can be assigned to any instrument, however we try to be mostly consistent. Common assignments for code letters are:

- A – aerosol data system active and logging enabled
- B – Best UPS active and logging enabled
- C – CN contamination control algorithm active (not logged)
- D – CCN active and logging
- F – Filters active and logging enabled
- G – GPS (at BRW ‘G’ is for second filter rack)
- H – Humidograph active and logging enabled
- L – PSAP active and logging enabled
- M – system monitor file logging (always present)
- N – nephelometer (reference/dry) active and logging enabled
- O – nephelometer (humidified/wet) active and logging enabled
- P – PID Box – this is box with T/RH and flow control
- R – sometimes an R is used for the PSAP (esp. if it’s a 3-w PSAP), or for second PID box
- T – TSI mass flow meter
- U - μMAC1050 A/D I/O subsystem active
- V - μMAC1050 A/D I/O subsystem active
- W – used to be for Watlow (predecessor to PID box) but now could be 3rd μMAC1050

3.7.3. Root Menu <enter> or <esc>

Pressing <Esc> or <enter> brings up the root menu, with the following choices (shortcut keys are in the left column). Use the arrow keys to highlight the desired action and then press <Enter>: (Not all stations have all of these options – it depends what instruments are deployed.)

	Online log entry	allows the user to make an entry in the system log file. Think of this as an on-line logbook, and use it liberally. Each entry can be at most 80 characters long, and each entry is recorded with a time stamp. Please make an entry for any deviations from normal operation, such as exceptions that you note on
--	------------------	--

		the daily checksheets. These logbook entries will be sent to the GMD aerosol group by email shortly after the data are sent to Boulder.
1	Client Options	controls some features of the client program.
2	Disk operations	allows the user to flush any buffered data to disk, and to change to a new set of data files
3	Save Window Setup	Saves layout of windows if you have a set up you like
4	Instant Messaging	Starts Instant Messenger
A	Activate the aerosol status window	
B	Activate the Best UPS status window	
C	Activate the CNC status window (shows current concentration)	
D	Activate the CCN status window	
F	Activate the filter status window	
G	Activate the GPS status window (Filter#2 at BRW)	
H	Activate the humidograph status window	
L (or R)	Activate the PSAP status window (L or R depends on your system, only one will work)	
M	Activate the Aerosol contamination window	
N	Activate the nephelometer status window (Typically this is the 'dry' or 'reference' neph.)	
O	Activate the second neph status window (Typically this is a 'wet' or 'humidified' neph.)	
P	Activate the PID box status window. (PID box controls T/RH/flow)	
U	Activate the umac status window (This is the main umac.)	
V	Activate the 2 nd umac status window (This is the umac that typically controls the filter carousel.)	
W	Activate the Watlow status window (The Watlow controls things like temperature and relative humidities.)	At Barrow W is the 3 rd umac box (associated with filter#2)
	Quit	

3.7.4. Aerosol Status Window and Monitor Menu <enter><A><M>

<p>The aerosol status window shows the values of monitoring parameters like temperatures, humidities, pressures flows and sampling size cut (1 or 10 um).</p>	<p>The aerosol menu allows you to disable logging, bypass the instruments, set the cut size, disable cut-size switching and change the reference voltage of the mass flow controllers.</p>

Window hide/Window unhide <1> changes visibility of status window

Analyzers bypass/Analyzers unbypass <2> context-sensitive function to bypass/unbypass the continuous analyzers

Cut coarse/Cut fine <3> context-sensitive function to toggle between sub 10 µm and sub 1 µm cut size

Disable cutpoint scanning <4> turns off switching between impactor size cuts

Analyzer flow <5> goes to menu for controlling flow thru instruments (e.g., mass flow controller). 1 volt corresponds to 10 lpm; the typical setting is 3 V to get 30 lpm. (This option is only used on aircraft systems.)

3.7.5. Nephelometer Status Window and Menu <enter><N><M> or <enter><O><M>

	<p>This window shows the current state of the nephelometer, including measured light scattering and back scattering at 3 wavelengths, neph monitoring parameters like temperature and voltage and neph state.</p>
<p>the neph menu allows the user to do a span check and change neph parameters.</p>	<p>The neph change parameters menu lets the user talk directly to the neph and affect its operating conditions.</p>

Window hide/Window unhide <1> changes visibility of status window

Span Check /Abort Span Check <2> context-sensitive menu to start/abort a CO₂ span check.


Change parameters (formerly 'Send command') <3> Activates a menu for sending selected commands to the nephelometer. Not normally used except for troubleshooting.

Start Neph zero <4> starts neph zero

Display counts <C> displays nephelometer photon counts

3.7.6. PSAP Menu <enter> <L> <M> or <enter><R> <M>

The CLAP window shows the light absorption flows, transmittance through the filter and various other monitoring parameters from the PSAP.



	Blue	Green	Red
N Bap (1/Mm):	2.10	2.20	1.80
B Bap(I) (1/Mm):	1.94	2.03	1.66
B Transmittance:	0.495	0.582	0.610
B Ir correction:	1.426	1.534	1.569
B I_sample:	46228462	83972905	39080796
B I_ref:	57587275	72179944	36450093
B I_sam rsd:	0.004008	0.004008	0.003999
B I_ref rsd:	0.003983	0.003986	0.003982
A I_normalized:	0.802755	1.163383	1.072173
A dI:	0.000014	0.000018	0.000024
N Flow (slpm):	1.13		



the CLAP menu is used to disable logging of absorption data and to note when a filter change takes place.

Window hide/Window unhide <1> changes visibility of status window

Start filter change/ End filter change <2> Context-sensitive menu to start/end a filter
Use this function to prevent data logging during a PSAP filter change, and to generate automatic log file entries when PSAP filters are changed.

3.7.7. Filters Menu <enter> <F> <M>

Window hide/Window unhide <1> changes visibility of status window

Filters bypass/Filters unbyypass <2> context sensitive function to allow manual bypassing of filter samples

Heater power <3> Set in watts

Q Filter Flow <4> goes to menu for controlling flow thru instruments (e.g., mass flow controller). 1 volt corresponds to 10 lpm; the typical setting is 3 V to get 30 lpm. (Only used on aircraft systems.)

Filter change <5> context-sensitive function to notify system that the filters will be changed. Changes to "Filter start" when changing.

NOTE: to cycle through filters to be sure they are all changing need to adjust a parameter in cpd.ini file. Talk to station scientist to figure this out.

3.7.8. Humidograph Menu <enter> <H><M>

Window hide/Window unhide <1> changes visibility of status window

Change Setpoint <2> displays submenu for changing setpoints manually

Set Channel A <A> Sets RH set point of reference nephelometer

Set Channel B Sets RH set point of humidified nephelometer

Set Channel C <C> Sets pump direction and speed. (range between -20 and 100)

0 = not pumping

15 = pumping water at 15% power into humidifier

-20 = pumping water backwards out of humidifier at 20% power

Set Channel D <D> Sets whether the high-range flow valve is open (H) or closed (L).

Disable/enable scanning <3> start or stop the humidifier from humidifying

3.7.9. CCN Menu <enter> <D><M>

Window hide/Window unhide <1> changes visibility of status window


Enable/disable setpoint scanning <2> toggle switch

Change delta T <4> enables user to change delta T

Display housekeeping info <H>

Display cumulative number distribution <N> makes a simple plot of CCN N(dp)

CCN housekeeping info	CCN cumulative distribution plot
-----------------------	----------------------------------



4. SAFETY

4.1. Heater Control

If the pump fails, any of the heaters (stack, humidifier rack inlet, preheater, humidifier, filter rack inlet) could overheat because the PID sensors don't have flow past them. There is a high-temperature cut out switch on the heaters, which limits the temperature on the outer tube to about 80 degrees Celsius.

4.2. Electrical Safety

We have insulated the electrical connections inside the rack, so that it is safe to perform standard tasks inside the rack (e.g., changing impactor films) with the power on. However for non-standard tasks (e.g., replacing the humidifier), you should always remove power from the rack and the instruments you are working on..

5. SPECIFICATIONS

5.1. Power Draw at 115 V DC

Equipment	Draw (A)	Surge (A)
Carbon vane pump	8.6	20
Diaphragm pump	2.4	12
Bypass blower	1.3	3
Cooling fan	2.7	3.5
Aerosol System	9.0	

6. ADDITIONAL COMPUTER INFO

6.0. Data flow

The 'cpd' server logs data in the /aer/stn/log directory, where "stn" is the identifier of the station. Every six hours (or 24 hours), the current data are archived in a zip file, for subsequent transmission to GMD in Boulder. The program that does the archiving is /aer/prg/cpd/cpd.archive, which is scheduled to run by the cron daemon. The configuration file for cron is /aer/prg/crontab/crontab.stn. The archived data are stored locally in the /aer/stn/log/sent directory. In Boulder, the data are initially put in the /ftp/incoming/aer/stn directory on the <ftp.cmdl.noaa.gov> server.

6.1. Aerosol Server Program (cpd)

Operation of cpd is controlled by a plain-text configuration file /aer/prg/cpd/etc/cpd.ini. Configuration information for different stations are contained in separate files; for example, the SPO configuration file is /aer/prg/cpd/etc/cpd.ini.spo. Normally, cpd.ini is a symbolic link file that points to the station-specific configuration file. Detailed information on the entries in cpd.ini can be found in /aer/prg/cpd/usr/doc/cpdini.txt.

The server program is designed to run as a daemon, which means that the user does not directly interact with the program. It is started with the 'cpd -d' command, and terminated with the 'cpd.kill' command. To verify that the server program is operating, type 'cpd.ps' from a command prompt. This will display the processes associated with the cpd program.

The latest version of the cpd software can be downloaded from our ftp site:

<ftp://ftp.cmdl.noaa.gov/aerosol/etc/cpd/cpd.iso>

Stations are upgrading to cpd2 which can be downloaded from

<ftp://ftp.cmdl.noaa.gov/aerosol/etc/cpd/livecpd2.iso>

6.2. User account

The computer is configured to automatically log into the 'cpd' user account, which is used for running the cpd program at field stations.

6.3. Location of this document

This document is stored in /aer/doc/ops_man/oper_man_gen_cpd.doc.

7. STATION SPECIFIC APPENDIX